

CONFIRMATION

Umweltbundesamt

**Announcement about the uniform practice in
monitoring emissions and ambient air,
circular from the Federal Environment Ministry (BMU) of 2007-04-12,
publication BAnz. 2007-04-20, no. 75, p. 4139 - 4140**

IV. Suitability of measuring equipment for the continuous monitoring of ambient air

Under reference to the number 3.2 of the announcement that for the realisation of the guideline 96/62/EG the advice's of the 27th September 1996 about the judgement and control of the air quality responsible places of the 1st October 1998 (BAnz. P. 15.126) the suitability is announced in the task of the BMU to the following measuring equipment:

4 Nitrogen oxides (NO, NO₂, NO_x).

4.1 AC32M

Manufacturer:

Environnement S.A., Poissy Cedex, Frankreich und
Ansyco GmbH, Karlsruhe, Deutschland

Suitability:

Continuous measurement of NO, NO₂ und NO_x in ambient air.

Measuring ranges during the suitability test:

NO₂ 0 - 400 µg/m³
NO₂ 0 - 500 µg/m³
NO 0 - 1200 µg/m³

Software version:

V2.45

Restriction:

-


Remark:

-


Test Report:

TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Köln, TÜV Rheinland Group
936/21205818/A vom 2006-12-08

Köln, 2007-05-04



Dr. P. Wilbring



Dipl.-Chem. M. Kerpa