

## Thermo Scientific Carbon Monoxide Analyzer, Model 48i

Gas filter correlation analyzer



### Key Features

- Approved to meet the following standards: US EPA, UK Environmental Agency and the European Union
- Ethernet connectivity for efficient remote access
- Enhanced user interface with one button programming and large display screen
- Flash memory for increased data storage and user downloadable software

The Thermo Scientific Carbon Monoxide (CO) Analyzer, Model 48i utilizes gas filter correlation technology to measure the amount of carbon monoxide in the air.

The Model 48i is based on the principle that carbon monoxide (CO) absorbs infrared radiation at a wavelength of 4.6 microns. Because infrared absorption is a nonlinear measurement technique, it is necessary for the instrument electronics to transform the basic analyzer signal into a linear output.

The Model 48i uses an exact calibration curve to accurately linearize the instrument output over any range up to a concentration of 10,000ppm.

This state-of-the-art gas analyzer offers features such as an ethernet port as well as flash memory for increased data storage.

Ethernet connectivity provides efficient remote access, allowing the user to download measurement information directly from the instrument without having to be on-site.

You can easily program short-cut keys to allow you to jump directly to frequently accessed functions, menus or screens. The larger interface screen can display up to five lines of measurement information while primary screen remains visible.

## Product Specifications

To maintain optimal product performance, you need immediate access to experts worldwide, as well as priority status when your air quality equipment needs repair or replacement. We offer comprehensive, flexible support solutions for all phases of the product life cycle. Through predictable, fixed-cost pricing, our services help protect the return on investment and total cost of ownership of your Thermo Scientific air quality products.

### Carbon Monoxide Analyzer, Model 48i

<b>Preset Ranges</b>	0-1, 2, 5, 10, 20, 50, 100, 200, 500, 1000, 2000, 5000 and 10000 ppm 0-1, 2, 5, 10, 20, 50, 100, 200, 500, 1000, 2000, 5000 and 10000 mg/m <sup>3</sup>
<b>Custom Ranges</b>	0-1 to 10000 ppm 0-1 to 10000 mg/m <sup>3</sup>
<b>Zero Noise</b>	0.02 ppm RMS (60 second averaging time)
<b>Lower Detectable Limit</b>	0.04 ppm
<b>Zero Drift (24 hour)</b>	< 0.1 ppm
<b>Span Drift (24 hour)</b>	+/-1% full scale
<b>Response Time</b>	60 seconds (30 second average time)
<b>Precision</b>	+/-0.1 ppm
<b>Linearity</b>	+/-1% full scale ≤ 1000 ppm +/-2.5% full scale > 1000 ppm
<b>Sample Flow Rate</b>	0.5-2 liters/min.
<b>Operating Temperature</b>	20°C - 30°C
<b>Power Requirements</b>	100 VAC, 115 VAC, 220-240 VAC +/-10% @ 275W
<b>Size and Weight</b>	16.75"(W) x 8.62"(H) x 23"(D), 49 lbs. (22.2 kg)
<b>Outputs</b>	Selectable Voltage, RS232/RS485, TCP/IP, 10 Status Relays and Power Fail Indication (standard) 0-20 or 4-20 mA Isolated Current Output (optional)
<b>Inputs</b>	16 Digital Inputs (standard), 8 0-10vdc Analog Inputs (optional)
<b>Approvals and Certifications</b>	US EPA Reference Method: RFCA-0981-054 MCERTS Certified: Sira MC070095/00 EN14626: TÜV 936/21203248/A Report

### Ordering Information

#### Carbon Monoxide Analyzer, Model 48i

Choose from the following configurations/options to customize your own Model 48i

#### 1. Voltage options:

A = 120 VAC 50/60 Hz (standard)  
B = 220 VAC 50/60 Hz  
J = 100 VAC 50/60 Hz

#### 2. Internal zero / span and/or Oxygen Sensor:

N = No zero / span valve assembly (standard)  
A = No zero/ span valve w/ Zero Air Scrubber  
Z = Internal zero / span valve assembly  
C = Internal zero / span valve w/ Zero Air Scrubber  
G = Oxygen Sensor with NO Zero/Span  
R = Oxygen Sensor with Zero/Span

#### 3. Filter wheel purge:

S = Standard plumbing (standard)  
P = Filter wheel purge setup

#### 4. Optional I/O:

A = None (standard)  
C = 4-20mA current output - 6 channels,  
0-10v analog input - 8 channel

#### 5. Mounting Hardware:

A = Bench mounting (standard)  
B = Ears & handles, EIA  
C = Ears & handles, retrofit

**Your Order Code: 48i -** \_ \_ \_ \_ \_

Lit\_48iAQI\_12/09

© 2009 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific Inc. and its subsidiaries.

This specification sheet is for informational purposes only and is subject to change without notice. Thermo Fisher Scientific makes no warranties, expressed or implied, in this product summary. Not all products are available in all countries. Please consult your local sales representative for details.

This product is manufactured in a plant whose quality management system is ISO 9001 certified.

#### Air Quality Instruments

27 Forge Parkway  
Franklin, MA 02038 USA

(866) 282-0430  
(508) 520-0430  
(508) 520-1460 fax

[www.thermo.com/air](http://www.thermo.com/air)

**Thermo**  
SCIENTIFIC